

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A & 2B , Raja S.C. Mullick Road, Jadavpur, Kolkata-700032

Corrigendum Notice

Ref: NIT No. IACS-G/ICAM/2023-24/58 Date: 31/01/2024

Please note the following additions/corrections:

Previous Version	Modified Version
X-ray Detector & Optics: X-ray Detector: Should operate in ----- by twin and long unit cell protein crystals. The active area should be, <u>at least 100 cm² or larger</u> with higher theta ----- for more accurate structures.	X-ray Detector & Optics: X-ray Detector: Should operate in ----- by twin and long unit cell protein crystals. The active area should be, <u>at least 80 cm² or larger</u> with higher theta ----- for more accurate structures..

All other terms & Conditions shall remain unchanged in the above-mentioned NIT.

Registrar